



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 3/1/14-CD

REVIEWED

REPORT #6

By Carol Day at 3:13 pm, Mar 19, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201186	NAME OF AGENCY ROLLA POLICE DEPARTMENT	DATE OF INSPECTION 02-25-2014
LOCATION OF INSTRUMENT (STREET AND CITY) 1007 N. ELM ST. ROLLA, MO. 65401		TIME OF INSPECTION 0853 HRS.

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02-25-2014 0853 HRS.
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER REPCO. MARKETING, INC. LOT # 13001 EXP. DATE 03-07-2015	
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34° °C SIMULATOR SN 502725 EXP. DATE 07-10-2014	
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	

TEST 1 .095	TEST 2 .096	TEST 3 .097
--------------------	--------------------	--------------------

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 5 (0-.04)	2 (.05-.09)	2 (.10-.14)	4 (.15-.19)	2 OVER .19	0
---------------------------	--------------------	--------------------	--------------------	-------------------	----------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE Kenneth L. Moberly	PRINT FULL NAME KENNETH L. MOBERLY
TYPE II PERMIT NUMBER/EXPIRATION DATE 220072 / 03-08-2014	TELEPHONE NUMBER (673) 308-1213

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

IN THE CIRCUIT COURT OF PHELPS COUNTY, MISSOURI
ASSOCIATE DIVISION

STATE OF MISSOURI, Plaintiff)
vs.) Case No. _____
_____, Defendant)

AFFIDAVIT PURSUANT TO SECTION 490.692 RSMo

Before me, the undersigned authority, personally appeared KENNETH L. MOBERLY, who, being by me duly sworn, deposes as follows: My name is KENNETH L. MOBERLY. I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am a "Type II Breathalyzer Operator," authorized by the Missouri Department of Health to carry out maintenance checks on a breathalyzer, and I am the custodian of the records of the breathalyzer for the Rolla Police Department, Rolla, Missouri. Attached hereto are _____ pages of records kept by the Rolla Police Department, Rolla, Missouri, in the regular course of business of the Rolla Police Department, Rolla, Missouri, for an employee or representative of the Rolla Police Department, Rolla, Missouri, with knowledge of the act, event, condition, opinion, or diagnosis recorded to make the records or to transmit information thereof to be included in such record; and the records was made at or near the time of the act, event, condition, opinion, or diagnosis. The records attached hereto are the original or exact duplicates of the original.

Kenneth L. Moberly
Affiant

In witness whereof I have hereunto subscribed my name and affixed my official seal this 25 day of February, 2014.

Jason D. Smith
Notary

My Commission Expires:

5-21-2017



JASON D. SMITH
My Commission Expires
May 21, 2017
Phelps County
Commission #13888487

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201186
02/25/14

TESTING OFFICER:
MOBERLY/KENNETH/L
OFFICER I.D.: 998
PERMIT NUMBER: 220072
EXPIRATION DATE: 03/08/14
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE

--- SUPERVISOR MODE ---

BLANK TEST	.000	08:57
INTERNAL STANDARD	VERIFIED	08:57
EXTERNAL STANDARD	.095	08:57
BLANK TEST	.000	08:58
EXTERNAL STANDARD	.096	08:59
BLANK TEST	.000	08:59
EXTERNAL STANDARD	.097	09:00
BLANK TEST	.000	09:01

N = 3
SIM. = .1
AVG. = .096

Operator Signature

Kenneth L. Moberly

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201186
02/25/14

ARREST TIME: 08:00
SUBJECT NAME:
MOBERLY/KENNETH/L
DOB: 11/26/63 SEX: M
STATE/D.L.: MO/123456789
ARRESTING OFFICER:
MOBERLY/KENNETH/L
OFFICER I.D.: 998
TESTING OFFICER:
MOBERLY/KENNETH/L
OFFICER I.D.: 998
PERMIT NUMBER: 220072
EXPIRATION DATE: 03/08/14
MISCELLANEOUS DATA:
MONTHLY RFI

--- BREATH ANALYSIS ---

BLANK TEST	.000	09:04
INTERNAL STANDARD	VERIFIED	09:04
RADIO INTERFERENCE		

Operator Signature

Kenneth L. Moberly

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201186
02/25/14
08:53

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~

Operator Signature

Kenneth L. Melrose

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



KENNETH L MOBERLY

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 03/08/2012

Number 220072

Expires 03/08/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)